

National Textiles & Garment Quality Supervision Inspection Centre (Zhejiang) Test Report

ADDRESS: No.300 XIASHA ROAD
HANGZHOU CHINA

Tel: (0571) 85122669

Fax: (0571) 85125121

Postcode: 310018

NUMBER:W202004486E

DATE: 2020-04-08

APPLICANT: Celecare Medical Wenzhou Co., Ltd

SAMPLE DESCRIPTION: KN95 MASK

ADDRESS: No2-No2 Building, Dapeng Road, Ouhai Area, Wenzhou

ITEM	STANDARD (KN95)	RESULT	RATING
1. FILTER EFFICIENCY (SALT MEDIUM) (GB 2626-2006)(%)			
-	≥ 95.0	96.4	PASS
2. RESPIRATORY RESISTANCE (GB 2626-2006)(Pa)			
-INSPIRATORY RESISTANCE	≤ 350	110	PASS
-EXPIRATORY RESISTANCE	≤ 250	101	PASS
PICTURE			

Remarks: The sample is not pretreated and the result is for reference.

俞杰

PREPARED AND CHECKED BY

YU JIE

----END----